y

Notice of Allowability	Application No.	Applicant(s)
	10/721,940	COPPOM ET AL.
	Examiner	Art Unit
	Distant L. Obissa	4704
	Richard L. Chiesa	1724
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. X This communication is responsive to the telephone interview conducted on December 12, 2005.		
2. The allowed claim(s) is/are <u>claims 1-24.</u>		
 3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) ☐ All b) ☐ Some* c) ☐ None of the: 1. ☐ Certified copies of the priority documents have been received. 		
2. Certified copies of the priority documents have been received in Application No		
3. Copies of the certified copies of the priority documents have been received in this national stage application from the		
International Bureau (PCT Rule 17.2(a)).		
* Certified copies not received:		
Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application. THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.		
4. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.		
5. CORRECTED DRAWINGS (as "replacement sheets") must be submitted.		
(a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached		
1) hereto or 2) to Paper No./Mail Date		
(b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date		
Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).		
6. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.		
Attachment(s) 1. Notice of References Cited (PTO-892)	5 Notice of Informal B	atent Application (PTO-152)
Notice of References Cled (PTO-092) Notice of Draftperson's Patent Drawing Review (PTO-948)	6. ☑ Interview Summary	· · · · · · · · · · · · · · · · · · ·
	Paper No./Mail Dat	e
Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date Symmon's Commont Regarding Requirement for Deposit	_	
 Examiner's Comment Regarding Requirement for Deposit of Biological Material 	_	nt of Reasons for Allowance
	9. Other	

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicants, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with applicants' attorney, Mr. Kent Lembke (Reg. No. 44,866) on December 12, 2005.

The application has been amended as follows:

The phrase --distributed over the entire surface of the filter media-- has been inserted between "locations" and the period at the end of claims 1 and 12.

The phrase --distributed over the entire surface of the filter media-- has been inserted between "locations" and the semi-colon in claim 17, line 6, and claim 24, line 10.

The word --fibrous-- has been inserted between "a" and "filter" (claim 23, line 2). The phrase "an electrode" (claim 23, line 6) has been changed to --a conductive electrode--. The phrase --in a plurality of substantially planar locations distributed over the entire surface of the filter media-- has been inserted between "media" and the semi-colon (claim 23, line 7).

2. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Richard L. Chiesa whose telephone number is (571) 272-1154.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Duane S. Smith, can be reached at (571) 272-1166.

Application/Control Number: 10/721,940

Art Unit: 1724

Any inquiry of a general nature or relating to the status of this application should be directed to the Technology Center 1700 receptionist whose telephone number is (571) 272-1700.

Page 3

Facsimile correspondence must be transmitted through (571) 273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Richard L. Chiesa December 16, 2005

> RICHARD L. CHIESA PRIMARY EXAMINER ART UNIT 1724

Richard L. Chiesa

Dec. 16, 2005